

Supported by: German Cancer Research Center University Hospital Carl Gustav Carus Dresden Carl Gustav Carus Faculty of Medicine, TU Dresden Helmholtz-Zentrum Dresden-Rossendorf

Core Unit for Molecular Tumor Diagnostics

Input Recomendation for selected Protocol

Protocol	Takara small RNA Seq	
Input Material	RNA	
Minimum Input	1,0 ng	
Standard Input	depends on Material	
Maximum Volume	6,0 μL	
Minimal Concentration	0,17 ng/μL	
Standard Concentration	depends on Material	
Extraction derived from	Non-FFPE	
RNA Qualityparameters		
RIN	depends on Material	
DV200	not considered	

For more Information please visit the manufacturers homepage

https://www.takarabio.com/products/next-generation-sequencing/epigenetics-and-small-rna-sequencing/small-rna-seq-kit

The CMTD is not responsible for the consequences/results, if an ordering work group wants to sequence RNA/DNA samples that failed the predefined quality parameters for sequencing!

Possible, risks which can occur are for example bad sequencing quality, no sequencing possible due to failure in library generation, introducing false signals of gene expression or allele frequency.

Erstellt: Alexander Krüger	Geändert: Alexander Krüger	Geprüft: Sandra Oster	Freigegeben:
Datum: 28.09.2021	Datum: 25.02.2022	Datum:	Datum:

Version 03 Seite 1 von 1